

# Surface performances of PVD ZrN coatings in biological environments

**C.E. RAMOUL, N.E. BELIARDOUH, R. BAHI, C. NOUVEAU, A. DJAHOUDI, M.J. WALOCK**

**Abstract:** Zirconium nitride (ZrN) thin films were deposited by reactive RF magnetron sputtering on Ti-6Al-4V and Si (100) substrates for potential use in biomedical applications. The tribological behaviour was evaluated against bovine bone in dry condition using a pin-on-disc apparatus. Abrasion is the primary wear mechanism observed in ZrN/bone contact. The corrosion properties were determined through two electrochemical techniques: potentiodynamic polarization and electrochemical impedance spectroscopy. The coatings with reduced oxygen content provided: (i) good resistance against corrosion when exposed to physiological solution and (ii) better anti-bioadhesion against *Staphylococcus aureus* bacteria.

**Keywords :** PVD coating, ZrN, wear, corrosion, bioadhesion